


Determination of molar masses via a measurement of the boiling point elevation (ebullioscopy) (Item No.: P3021900)

Curricular Relevance



Difficulty



Difficult

Preparation Time



10 Minutes

Execution Time



20 Minutes

Recommended Group Size



2 Students

Additional Requirements:

- vice
- solvents as required
- Precision balance, 620 g / 0.001 g

Experiment Variations:

Keywords:

molar mass, boiling point elevation, ebullioscopy, ebullioscopic constant

Overview

Short description

Principle

Didactic setup to train and demonstrate the determination of molar masses by way of a measurement of the boiling point elevation. The boiling point elevation of aqueous solutions of different substances is determined using. The ebullioscopic constant of water is calculated from the experimental results.



Fig. 1

Safety information

If solvents other than water are used for the experiment, ensure proper handling and disposal.

Safety instructions



Hydroquinone

H302: Harmful if swallowed.

H317: May cause an allergic skin reaction.

H351: Suspected of causing cancer.

P280: Wear protective gloves/protective clothing/eye protection/face protection.

Equipment

Position No.	Material	Order No.	Quantity
1	Retort stand, h = 750 mm	37694-00	1
2	Right angle boss-head clamp	37697-00	3
3	Universal clamp	37715-00	3
4	Apparatus for elevation of boiling point	36820-00	1
5	Temperature meter digital, 4-2	13617-93	1
6	Temperature probe, immersion type, Pt100	11759-01	1
7	Protective sleeves f.temp.probe,2	11762-05	1
8	Flask,round,1-neck,250ml,GL25/13	35812-15	1
9	Beaker, high, BORO 3.3, 250 ml	46027-00	1
10	Gasket for GL25, 8mm hole, 10 pcs	41242-03	1
11	Silicone tubing i.d. 7mm, 1 m	39296-00	1
12	Mortar with pestle, 150 ml, porcelain	32604-00	1
13	Pinchcock, width 15 mm	43631-15	1
14	Microspoon, steel	33393-00	1
15	Wash bottle, plastic, 500 ml	33931-00	1
16	Pellet press for calorimeter	04403-04	1
17	Heating mantle f. roundbottom flask, 250ml	49542-93	1
18	Clamp for heating mantle	49557-01	1
19	Power regulator	32288-93	1
20	Weighing dishes, square shape, 84 x 84 x 24 mm, 25 pcs.	45019-25	1
21	Funnel, glass, top dia. 80 mm	34459-00	1
22	Pasteur pipettes, 250 pcs	36590-00	1
23	Rubber caps, 10 pcs	39275-03	1
24	Boiling beads, 200 g	36937-20	1
25	Desiccator, vacuum, diam. 150 mm	34126-00	1
26	Porcelain plate f.desiccator150mm	32474-00	1
27	Watch glass, dia.80mm	34572-00	1
28	Glycerol, 250 ml	30084-25	1
29	Urea, 250 g	30086-25	1
30	Hydroquinone 250 g	30089-25	1
31	Water, distilled 5 l	31246-81	1

Tasks

1. Determine the boiling point elevation of aqueous solutions of different substances
2. Calculate the ebullioscopic constant of water from the experimental results

Setup and procedure

Setup



Measurement principle




Fig. 2

The measurement solution (solution or solvent) is in the inner vessel. Bring some solvent in a round bottom flask to a boil. The solvent vapour flows first through the outer vessel and then through the inner vessel. In doing so, it heats the solution that is located in the inner vessel.

Setup

The apparatus for measuring the boiling point elevation consists of two cylindrical glass vessels that are placed with one inside the other. At its bottom and just below the GL 45 thread, the outer vessel is equipped with a glass tube connector ($d = 8 \text{ mm}$).

Apart from a GL 25 threaded connector and a lateral glass tube connector ($d = 8 \text{ mm}$), the inner vessel has a narrow glass tube inside. The inlet opening of this glass tube commences in the wall of the vessel. When assembling the two vessels, the inlet opening must be below the gasket (see Fig. 2). The gasket and screw cap ensure the gas-tight connection of the two vessels. Ensure that the opening (a) is not sealed by the gasket! Prior to assembling and screwing the two vessels together, determine the exact mass of the dry inner vessel by weighing it. Note down the result ($=m_1$). Connect the assembled apparatus via the lower glass tube connector of the outer vessel to the 250 ml round bottom flask. Tighten the screw connection tightly. Prior to doing so, replace the gasket of the screw connection cap with a gasket with an 8 mm hole. Fill the flask with 150 to 200 ml of the solvent to be used as well as with several boiling beads. Place the flask in a heating mantle.

Connect short pieces of silicone tubing to the lateral glass tube connectors. If water is used as the solvent, lay the tubings simply into a beaker but ensure that they do not reach down to the bottom. Instead, they should end approximately in the middle of the beaker. If organic solvents are used, connect these tubings via a Y-shaped connecting tube to a double-neck flask (100 ml) that is

equipped with a reflux condenser (Dimroth condenser). Attach a pinchcock to the tubing that is connected to the outer vessel but leave the pinchcock open for the time being.

Fill approximately 40 ml of solvent into the inner vessel and seal the vessel with a protective sleeve for the temperature sensor. To do so, replace the gasket of the screw connector with the gasket with the 12 mm hole that was removed from the round bottom flask beforehand. Apply two to three drops of glycerol to the protective sleeve for better heat transfer and insert the temperature probe. Connect the temperature probe to the digital temperature meter.

The substance whose molar mass is to be determined must be provided in the form of pellets. A simple pellet press can be used for the production of the pellets (Fig. 3). It works as follows:

Grind the substance finely in a mortar. Place the small steel die into the hole of the press cylinder. It seals the bottom of the hole. Then, fill the finely ground substance into the hole. Place the longer die from the top onto the hole, thereby compressing the substance slightly. After that, clamp the press into a vice as shown in Fig. 3, thereby pressing against the dies so that the substance is pressed into a compressed pellet. After the pressing process, push the pellet out of the cylinder by way of the long die. The measurement requires 1 to 2 pellets with a thickness of approximately 5 to 7 mm. Determine their mass (m_s) by weighing with an accuracy of at least 1 mg.




Fig. 3

Procedure

Measurement

Heat the solvent in the flask up to its boiling temperature. The resulting vapour rises into the outer vessel and heats the inner vessel. In doing so, part of the vapour condenses; the other part escapes via the tubing with the pinchcock into the beaker or into the flask with a reflux cooler. After a few minutes of boiling, lower the heating mantle for a few seconds so that the boiling stops briefly and the solvent that has condensed in the outer vessel flows back into the flask. When this has happened, raise the heating mantle once again against the flask and continue boiling. When vapour rises once again, close the pinchcock so that the vapour now flows via the inner glass tube into the inner vessel and there through the solvent. Once the boiling temperature is reached and remains constant, switch the temperature meter to the mode for measuring the change in temperature with a resolution of one-hundredth of one degree (button $\Delta T/K$; please refer to the operating instructions of the measuring instrument). Open the inner vessel, lift the protective sleeve with the temperature probe up, add a substance pellet to the solvent as quickly as possible, and close the apparatus again. Now, the boiling point of the solution in the inner vessel is determined.

The temperature meter indicates directly the difference with regard to the boiling temperature of the pure solvent (ΔT). Then, open the pinchcock (Important! This is to ensure that the liquid in the inner vessel cannot be sucked into the flask when the heating apparatus is switched off.). It is only then that the heating mantle can be switched off and lowered so that the boiling in the flask stops quickly. In order to determine the mass of the solvent ($= m_S$), remove the thermometer. Remove the inner vessel carefully (Attention! It is still very hot.) and weigh it together with the solution inside. Note down the mass thus determined ($= m_2$).

Theory and evaluation

Evaluation

The following values were measured:

- Mass of the empty vessel m_1
- Mass of the vessel containing the solution m_2
- Mass of the dissolved substance m_s
- Mass of the solvent $m_S = m_2 - m_1 - m_s$
- Boiling temperature of the solvent T_1
- Boiling temperature of the solution T_2
- Boiling point elevation $\Delta T = T_2 - T_1$

Since the boiling point elevation is proportional to the molar concentration (= mol on 1000 g of solvent), the following applies:

$$M = (ms - 1000 - K) / (mS - \Delta T)$$

M = molar mass

K = ebullioscopic constant (specific for every individual solvent)

It can be found in the literature.

Example of a measurement (urea/water):

$$ms = 0.685 \text{ g}$$

$$\Delta T = 0.13 \text{ K}$$

$$mS = 45.0 \text{ g}$$

$$K = 0.515 \text{ g/mol} \cdot \text{K}$$

(for water as the solvent)

$$M = \frac{0.685 \text{ g} \cdot 1000 - 0.515 \text{ g/mol} \cdot \text{K}}{45.0 \text{ g} - 0.13 \text{ K}} = 60.35 \text{ g/mol (Urea = 60.05 g)}$$

If the ebullioscopic constants of solvents are determined based on known substances, K results from:

$$K = \frac{M \cdot \Delta T \cdot mL}{ms \cdot 1000}$$

Notes

In order to train and demonstrate the determination of molar masses by way of a measurement of the boiling point elevation, water is absolutely sufficient as the solvent and urea or hydroquinone as the test substance. Of course, it must be ensured that the solid substances are dry when they are used. Dry them for at least 24 hours in a desiccator. When using other substances, determine the most suitable solvent that can be used. The ebullioscopic constants that are specific for every individual solvent can be found in the literature.

If other solvents are used in the place of water, do not let the resulting solvent vapour flow into a beaker. Instead, guide them via a V-connector into a flask with a reflux cooler (Fig. 1). For this purpose, the following additional equipment is required:

1 Round bottom flask, 100 ml, 2 necks, GL 25/12, GL 18/8	35842-15
1 Condenser, Dimroth type, GL 25/12	35816-15
2 Hose clips, d = 8...12 mm, 2 pcs.	40996-00
1 Connecting tubes, V-shaped, d = 8 mm	37014-00
1 Retort stand, h = 750 mm	37694-00
2 Boss head	37697-00
2 Universal clamp	37715-00